Data Logger for Cloud Storage

TR-7wb/nw Series Features and Specs


Measurement Items
Temperature
Humidity

Data Collection
Bluetooth®,
LAN / USB Connection

Data Access
T&D WebStorage Service,
Intranet, Local PC

Warning Notification
E-mail

TR-7wb/nw, with multiple types of communication interface (Bluetooth®, Wireless/Wired LAN and USB) have been designed to meet your temperature and humidity data management environment and needs.



- * T&D WebStorage Service is a free cloud-based storage service provided by T&D Corporation. A LAN based environment with Internet connection is necessary to use this service.
- * The Bluetooth® trademark and logo are registered trademarks owned by Bluetooth SIG, Inc. T&D Corporation uses these marks under license.

1 Upload via Bluetooth wb

Use our "T&D Thermo" app (Android or iOS) with Bluetooth to make settings, view recorded data in Graphs and share data to the cloud.

2 Auto-Upload to Cloud via LAN

Recorded data can be automatically uploaded to our T&D WebStorage Service via wireless wb or wired LAN nw. This allows for the accessing of recorded data from any smartphone or PC with an Internet connection.

 A LAN environment with an Internet connection is required.

3 Simple and Safe USB Communication

By simply connecting a logger to PC via USB, it is possible to download recorded data and view in Graphs or Lists.

* The required free-of-charge software is available for download.

Model	Measurement Items	Measurement Range	Notes
TR-71wb / nw	Temperature 2ch	-60 to 155 °C	The measurement range depends on the sensor type. Wide selection of optional sensors available
TR-72wb / nw	Temperature / Humidity 1ch Each	0 to 55 °C / 10 to 95%RH	
TR-72wb-S / nw-S	Temperature / Humidity 1ch Each	-25 to 70 °C / 0 to 99%RH	The supplied sensor for the S model provides higher accuracy to $\pm 2.5\% RH$
TR-75wb / nw	Temperature 2ch (Thermocouple)	-199 to 1760 °C	For use with Thermocouple Sensor Types: K, J, T, E, S, R

Sending Warning Report Mails

Warning e-mails can be sent upon T&D WebStorage Service receiving warning information from the data logger.

1.5 years of Operation on Two Batteries

Battery operation is possible for up to 1.5 years with just two AA alkaline batteries.

Large Logging Capacity: 8000 Readings per Channel

It is possible to record up to 8000 data readings in each of the two channels. If set at a recording interval of 60 minutes, it gives the user one year's worth of measurements.

Easy Operation via Front Buttons

It is possible to start and stop recording, change recording interval, and make auto-upload setting from the buttons on the face of the logger.



TR-7wb/nw Series - Specifications

		TR-71wb / 71nw	TR-72	2wb / 72nw	TR-72wb	-S / 72nw-S	TR-75wb / 75nw		
Measurement Channels		Temperature 2ch	Temperature 1ch, Humidity 1ch		Temperature 1ch, Humidity 1ch (High-Precision Type)		Temperature 2ch		
Sens	sor	Thermistor	Thermistor	Polymer Resistance	Thermistor	Polymer Resistance	Thermocouple: Type K, J, T, E, S, R (*1)		
	surement Units	°C, °F	°C, °F	%RH	°C, °F	%RH	°C, °F		
Me	Internal Sensor	-10 to 60 °C (*2)	-	-	-	-	-		
	External Sen- sor	-40 to 110°C (Supplied Sensor) -60 to 155°C (Optional Sensor)	0 to 55 °C	10 to 95 %RH	-25 to 70 °C	0 to 99 %RH (*3)	K -199 to 1370 °C E -199 to 1000 °C J -199 to 1200 °C S -50 to 1760 °C T -199 to 400 °C R -50 to 1760 °C		
Accı	uracy	Avg. ±0.3°C at -20 to 80 °C Avg. ±0.5°C at -40 to -20 °C 80 to 110 °C	±0.5°C	±5 %RH at 25°C, 50%RH	±0.3°C at 10 to 40 °C ±0.5°C all other tem- peratures	±2.5 %RH at 15 to 35 °C, 30 to 80 %RH	Thermocouple Measurement (Sensor inaccuracies not included) Type K, J, T, E: ±(0.5°C+0.3% of reading) at -100°C or above Type S, R: ±(1.5°C+0.3% of reading) at 100°C or above Cold Junction Compensation ±0.5°C at 10 to 40°C ±0.8°C other temperatures within the operating environment of the logger		
Meas lution	surement Reso- n	0.1 °C	0.1°C	1 %RH	0.1 °C	0.1 %RH	Type K, J, T, E: 0.1 °C Type S, R: approx. 0.2 °C		
Resp	oonsiveness	Thermal Time Constant: Approx. 75 sec. Response Time (90%): Approx. 190 sec.	Response Time (90%): Approx. 7 min.		Response Time (90%): Approx. 7 min.		-		
_CD [Display Items	Measurements (fixed or altern	nating display)	, Battery Warning Mar	k, etc.				
	ng Capacity	8,000 data sets (One data set				nit.)			
	rding Interval	Select from 15 choices: 1, 2, 5, 10, 15, 20, 30 sec. or 1, 2, 5, 10, 15, 20, 30, 60 min.							
Recor	rding Mode	Endless (Overwrite oldest data when capacity is full) or One Time (Stop recording when capacity is full)							
uto-ı	upload Interval	Select from 15 choices: OFF	No auto-uploa	ad), 1, 2, 5, 10, 15, 20,	30 min. or 1, 2, 3,	4, 6, 12, 24 hrs.			
Communication Interfaces		Protocol: HTTP(*5), DHCP, DNS TR-7wb: Bluetooth *Communication Bluetooth 4.2 (Bluetooth low energy) TR-7nw: Wired LAN Communication 100BASE-TX/10BASE-T (RJ45 Connector) Protocol: HTTP (*5), DHCP, DNS USB Communication : USB 2.0 (Mini-B connector)							
Powe	er (*6)	Battery: AA Alkaline x 2, AA Ni-MH x 2 External: USB Bus 5V 200mA, AC Adaptor AD-05A2 or AD-05C2, PoE IEEE 802.3af (TR-7nw only)							
Batt	ery Life (*7)	TR-71wb/TR-72wb: Approx. 10 days (when Auto-Approx. 1 year (when Auto-Approx. 15 months (when A*1.2 times longer with Bluet *Approx. 1.5 yrs with Blueto TR-71nw/TR-72nw: Approx. 10 days (when Auto-	upload interva uto-upload in ooth OFF ooth & Auto-Up	al is 1 hr) terval is 12 hr or more bload OFF)		TR-75wb: Approx. 10 days (when Auto-upload interval is 1 min) Approx. 10 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more) *1.2 times longer with Bluetooth OFF *Approx. 15 months with Bluetooth & Auto-Upload OFF TR-75nw:		
		Approx. 1 year (when Auto- Approx. 1.5 years (when Au 'Approx. 1.5 yrs with Auto-l	upload interva	al is 1 hr)			Approx. 10 days (when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more) 'Approx. 1 yr with Auto-Upload OFF		
	ensions	Approx. 1 year (when Auto- Approx. 1.5 years (when Au	upload interva	al is 1 hr)			(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more)		
Dime		Approx. 1 year (when Auto- Approx. 1.5 years (when Au *Approx. 1.5 yrs with Auto-l	upload interva	al is 1 hr)			(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more)		
Dime Weigl Opera	ht ating Environ-	Approx. 1 year (when Auto-Approx. 1.5 years (when Au*Approx. 1.5 yrs with Auto-I H 58 mm x W 78 mm x D 26 mm Approx. 55 g Temperature –10 to 60°C / –1 Humidity 90 %RH or less (no	upload intervato-upload intervato-upload inte Jpload OFF	al is 1 hr) rrval is 12 hr or more) rn using external powe	, ,,,		(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more)		
Dime Weigl Opera ment	ht ating Environ-	Approx. 1 year (when Auto-Approx. 1.5 years (when Au*Approx. 1.5 yrs with Auto-I H 58 mm x W 78 mm x D 26 mm Approx. 55 g Temperature –10 to 60°C / –1 Humidity 90 %RH or less (no Temperature Sensor TR-0106 x2	upload intervato-upload inte to-upload inte Jpload OFF 0 to 45°C whe condensation Temperatu	al is 1 hr) rrval is 12 hr or more) rn using external powe 1) re-Humidity Sensor HA-3001 x1	High Precision T	emperature-Humidi- SHA-3151 x1	(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more) *Approx. 1 yr with Auto-Upload OFF		
Dime Weigl Dperament Acces	ht ating Environ- ssories vare Compat-	Approx. 1 year (when Auto-Approx. 1.5 years (when Au*Approx. 1.5 yrs with Auto-I H 58 mm x W 78 mm x D 26 mm Approx. 55 g Temperature -10 to 60°C / -1 Humidity 90 %RH or less (no	upload intervato-upload intervato-upload inte Jpload OFF 0 to 45°C whe condensation Temperatu Ti Registration C 0 Graph, T&D I 64 bit, Microst	an using external powers) re-Humidity Sensor HA-3001 x1 Data Server (For PC)	High Precision T ty Sensor B Cable US-15C,	SHA-3151 x1 Manual Set (Warranty	(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more) *Approx. 1 yr with Auto-Upload OFF		
Dime Weigl Dpera ment Acces	ht ating Environ- ssories	Approx. 1 year (when Auto-Approx. 1.5 years (when Auto-Approx. 1.5 yrs with Auto-IIIIIIIIIIIIIIIIIIIIIIIIIIIIIIIIIIII	upload intervato-upload intervato-upload intervato-upload intervators of the second of	al is 1 hr) real is 12 hr or more) real is 12 hr or more) real using external power) re-Humidity Sensor HA-3001 x1 ode Label, USB Mini- Data Server (For PC) oft Windows 8 32 / 64	High Precision T ty Sensor B Cable US-15C, bit, Microsoft Wir	SHA-3151 x1 Manual Set (Warranty	(when Auto-upload interval is 1 min) Approx. 7 months (when Auto-upload interval is 1 hr) Approx. 1 year (when Auto-upload interval is 12 hr or more) *Approx. 1 yr with Auto-Upload OFF		

- *1: Compatible wire sizes are as follows.
 Single Wire: \$\phi\$ 0.32 to \$\phi\$ 0.65 mm (AWG 28 22), Twisted Wire: 0.08 to 0.32 mm² (AWG 28 22), \$\phi\$ 0.12 mm or more in diameter, Stripping Length: 9 to 10 mm
 *2: When Auto Upload is used frequently, the measurement of the internal sensor may rise by around 0.3°C. When using external power, the data logger itself generates heat and the internal sensor will report a temperature much higher than ambient; we recommend using an external temperature sensor in this case.
 *3 When continually used in environments with temperatures above 60 °C, accuracy of humidity measurements will decrease over time. Also, humidity cannot be measured at temperatures below -20 °C.

- tures below -20 °C.

 *4: If you wish to use the WPS feature, set the security type of the wireless LAN access point to "WPA2-PSK(AES)" or "None".

 *5: HTTP client. Proxy supported.

 *6: When using external power, the internal temperature of the logger rises.

 *7: Battery life varies depending upon multiple factors including frequency of communication, LAN environment, ambient temperature, recording interval, and battery performance.

 *All estimates are based on operations carried out with a new battery and are in no way a guarantee of actual battery life.

 *8: For installation, it is necessary to have Administrator (Computer Administrator) rights.

 *9: We recommend using an operating system in the same language as the display language. Operation in different languages is not guaranteed.

 The specifications listed above are subject to change without notice.

